

stomdevice

A-dec DV Dry Vacuum Quick Reference




Electrical Requirements

DV5	DV7	DV10	DV12
Required Circuit Breaker Size			
20 Amp	30 Amp	30 Amp	30 Amp
Hard Wired Connection (Blue/Brown/Green/Green)			
Maximum Current Draw			
14.1 Amp	20.6 Amp	22 Amp	24 Amp




CAUTION Local regulation may require licensed plumbers and electricians to install utilities. All plumbing and utilities must conform to prevailing local codes.

Below Grade Installation



Overhead Installation



A-dec Headquarters
2601 Crestview Drive
Newberg, Oregon 97132
United States
Tel: 1.800.547.1883 within USA/CAN
Tel: +1.503.538.7478 outside USA/CAN
Fax: 1.503.538.0276
www.a-dec.com

A-dec St. Louis
1601 Manufacturers Drive
Fenton, MO 63026 USA
Tel: 1.888.665.8607 within USA/CAN
www.a-dec.com

For more information, including trunk and piping guidelines and the installation of dual vacuums, scan the appropriate QR code to access the *A-dec DV Dry Vacuum Pre-Installation Guide* (p/n 86.0896.00) and the *A-dec DV Dry Vacuum Installation Guide* (p/n 86.0891.00) available in the Resource Center at www.a-dec.com.

Pre-Installation Guide

Installation Guide

86.0893.00 Rev B
Date of Issue 2021-09-29
Copyright 2021 A-dec, Inc.
All rights reserved

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ