stomdevice

EN

AMANNGIRRBACH



IMPLANT PROSTHETICS

ceramill®

MPS
Multi Platform Systems
by

MEDENTIKA®
A Straumann Group Brand

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

Content

4	**B-Series**	compatible with	Bredent Medical® | SKY®
10	**BS-Series**	compatible with	BEGO Implant Systems® | Semados® S-Line Semados® RS/RSX-Line | Semados® RI-Line
14	**C-Series**	compatible with	Altatec® | Camlog Screw-Line - Root-Line 2®
20	**CX-Series**	compatible with	Medentis Medical® | ICX®
24	**D-Series**	compatible with	Altatec® | Conelog®
30	**E-Series**	compatible with	Nobel Biocare® | NobelReplace® Tapered
36	**EV-Series**	compatible with	Dentsply® Implants | ASTRA TECH OsseoSpeed® EV
42	**F-Series**	compatible with	Nobel Biocare® | Nobel Active® NobelReplace® Tapered
48	**H-Series**	compatible with	Biomet 3i® | Certain®
54	**I-Series**	compatible with	Biomet 3i® | External Hex®
60	**K-Series**	compatible with	Nobel Biocare® | Brånemark®
66	**L-Series**	compatible with	Straumann® | Bone Level®
72	**N-Series**	compatible with	Straumann® | Tissue Level®
78	**R-Series**	compatible with	Zimmer Dental® | MIS® | BioHorizons®
86	**S-Series**	compatible with	DENTSPLY Implants® ASTRA TECH OsseoSpeed® TX
92	**T-Series**	compatible with	DENTSPLY Implants® | XiVE® S
98	**Y-Series**	compatible with	DENTSPLY Implants® | ANKYLOS®
104	**Tools**		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

B-SERIES



The B-Series abutments are compatible with

Bredent Medical®
SKY®

DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	D 3,5/4,0
Article No.	B 50

Abutment screw

- Titanium grade 5 KV
- Torx T6



Article No.	B 60
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	B 60 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5
Chimney height	3.5 mm
Gingiva height	0,3 mm
Article No.	B 1000
Abutment screw Laboratory screw	B 60 B 60 L
Please note:	This is used with Scanbody 2. generation.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5
Chimney height	5.5 mm
Gingiva height	0,3 mm
Article No.	B 1100
Abutment screw Laboratory screw	B 60 B 60 L
Please note:	This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5
Chimney height	3.5 mm
Gingiva height	0,3 mm
Article No.	B 1500
Abutment screw Laboratory screw	B 60 B 60 L
Please note:	This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5
Diameter	11.5 mm
Article No.	B 9000-AG
Abutment screw Laboratory screw	B 60 B 60 L
Please note:	This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,5/4,0
Article No.	B 1400
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 4,0	D 4,0	D 4,0	D 4,0	D 4,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	B 4000	B 4100	B 4200	B 4300	B 4400

MedentiBASE® prosthetic components

- B 4840 MedentiBASE® titanium base incl. bridge screw
- B 7600 MedentiBASE® bridge screw Torx T6
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	B 4840	B 7600

MedentiBASE® adhesive cap for passive-fit

- B 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- B 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	B 4800	B 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	B 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

BS-SERIES



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The BS-Series abutments are compatible with

BEGO Implant Systems®
Semados® S-Line
Semados® RS/RSX-Line
Semados® RI-Line

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	BS 60
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	BS 60 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 4,1	D 4,5	D 5,5
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,1 mm	0,1 mm	0,1 mm
Article No.	BS 1010	BS 1020	BS 1030
Abutment screw Laboratory screw	BS 60 BS 60 L	BS 60 BS 60 L	BS 60 BS 60 L
Please note:	This is used with Scanbody 2. generation.		

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 4,1	D 4,5	D 5,5
Chimney height	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,1 mm	0,1 mm	0,1 mm
Article No.	BS 1110	BS 1120	BS 1130
Abutment screw Laboratory screw	BS 60 BS 60 L	BS 60 BS 60 L	BS 60 BS 60 L
Please note:	This is used with Scanbody 2. generation.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm




Implant connection		D 4,1	D 4,5
Chimney height		3.5 mm	3.5 mm
Gingiva height		0,15 mm	0,15 mm
Article No.		BS 1510	BS 1520
Abutment screw Laboratory screw		BS 60 BS 60 L	BS 60 BS 60 L
Please note:	This is used with Scanbody 2. generation.		

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection		D 3,25-5,5
Diameter		11.5 mm
Article No.		BS 9000-AG
Abutment screw Laboratory screw		BS 60 BS 60 L
Please note:	This is used with Scanbody 2. generation.	

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection		D 3,25-5,5
Article No.		BS 1400
Please note:	The Scanbody is sterilisable and for intra-oral scanning.	

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» C-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The C-Series abutments are compatible with

Altatec®
Camlog Screw-Line – Root-Line 2®

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0
Article No.	C 50	C 51	C 52	C 53

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	C 60	C 61	C 62
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.		

Laboratory screw

- red coated



Article No.	C 61 L	C 62 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.	

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0	D 6,0
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm	0,3 mm	0,3 mm
Article No.	C 1000	C 1010	C 1020	C 1030	C 1040
Abutment screw	C 61	C 61	C 61	C 62	C 62
Laboratory screw	C 61 L	C 61 L	C 61 L	C 62 L	C 62 L
Please note:	This is used with Scanbody 2. generation.				

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0	D 6,0
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm	0,3 mm	0,3 mm
Article No.	C 1100	C 1110	C 1120	C 1130	C 1140
Abutment screw Laboratory screw	C 61 C 61 L	C 61 C 61 L	C 61 C 61 L	C 62 C 62 L	C 62 C 62 L
Please note:	This is used with Scanbody 2. generation.				

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,8	D 4,3
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	C 1500	C 1510	C 1520
Abutment	C 60	C 60	C 60
Please note:	This is used with Scanbody 2. generation.		

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0
Diameter	11.5 mm	11.5 mm	11.5 mm	11.5 mm
Article No.	C 9000-AG	C 9010-AG	C 9020-AG	C 9030-AG
Abutment screw Laboratory screw	C 61 C 61 L	C 61 C 61 L	C 61 C 61 L	C 62 C 62 L
Please note:	This is used with Scanbody 2. generation.			

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0/6,0
Article No.	C 1400	C 1410	C 1420	C 1430
Please note:	The Scanbody is sterilisable and for intra-oral scanning.			

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,3	D 3,3	D 3,3	D 3,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	C 4000	C 4100	C 4200	C 4300	C 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 30 Ncm



Implant connection	D 3,8	D 3,8	D 3,8	D 3,8	D 3,8
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	C 4010	C 4110	C 4210	C 4310	C 4410

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 30 Ncm



Implant connection	D 4,3	D 4,3	D 4,3	D 4,3	D 4,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	C 4020	C 4120	C 4220	C 4320	C 4420

MedentiBASE® prosthetic components

- C 4840 MedentiBASE® titanium base incl. bridge screw
- C 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	C 4840	C 7600

MedentiBASE® adhesive cap for passive-fit

- C 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- C 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	C 4800	C 4820

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	C 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

CX-SERIES



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The CX-Series abutments are compatible with

Medentis Medical®
ICX®

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Abutment screw

- Titanium grade 5 KV
- Hex 1.4



Article No.	CX 60
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	CX 60 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection		D 3,75-4,8	D 3,75-4,8
Chimney height		3.5 mm	3.5 mm
Gingiva height		0,65 mm	1,15 mm
Article No.		CX 1000	CX 1200
Abutment screw Laboratory screw		CX 60 CX 60 L	CX 60 CX 60 L
Please note:	This is used with Scanbody 2. generation.		

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection		D 3,75-4,8	D 3,75-4,8
Chimney height		5.5 mm	5.5 mm
Gingiva height		0,65 mm	1,15 mm
Article No.		CX 1100	CX 1300
Abutment screw Laboratory screw		CX 60 CX 60 L	CX 60 CX 60 L
Please note:	This is used with Scanbody 2. generation.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,75-4,8
Chimney height	3.5 mm
Gingiva height	0,65 mm
Article No.	CX 1500
Abutment screw Laboratory screw	CX 60 CX 60 L
Please note:	This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,75-4,8
Diameter	11.5 mm
Article No.	CX 9000-AG
Abutment screw Laboratory screw	CX 60 CX 60 L
Please note:	This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,75-4,8
Article No.	CX 1400
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

D-SERIES



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

The D-Series abutments are compatible with

Altatec®
Conelog®


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over cam

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Stainless steel



Implant connection	D 3,3	D 3,8/4,3
Article No.	D 50	D 51

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	D 60
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	D 60 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,3	D 3,8/4,3	D 3,8/4,3
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	D 1000	D 1200	D 1010	D 1210
Abutment screw Laboratory screw	D 60 D 60 L	D 60 D 60 L	D 60 D 60 L	D 60 D 60 L
Please note:	This is used with Scanbody 2. generation.			

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,3	D 3,8/4,3	D 3,8/4,3
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	D 1100	D 1300	D 1110	D 1310
Abutment screw Laboratory screw	D 60 D 60 L	D 60 D 60 L	D 60 D 60 L	D 60 D 60 L
Please note:	This is used with Scanbody 2. generation.			

Titanium base Bridges/ bars 2nd Generation


- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection		D 3,3	D 3,8/4,3
Chimney height		3.5 mm	3.5 mm
Gingiva height		0,65 mm	0,65 mm
Article No.		D 1500	D 1510
Abutment screw Laboratory screw		D 60 D 60 L	D 60 D 60 L
Please note:	This is used with Scanbody 2. generation.		

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection		D 3,3	D 3,8/4,3
Diameter		11.5 mm	11.5 mm
Article No.		D 9000-AG	D 9010-AG
Abutment screw Laboratory screw		D 60 D 60 L	D 60 D 60 L
Please note:	This is used with Scanbody 2. generation.		

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection		D 3,3	D 3,8/4,3
Article No.		D 1400	D 1410
Please note:	The Scanbody is sterilisable and for intra-oral scanning.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,3	D 3,3	D 3,3	D 3,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	D 4000	D 4100	D 4200	D 4300	D 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 3,8/4,3	D 3,8/4,3	D 3,8/4,3	D 3,8/4,3	D 3,8/4,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	D 4010	D 4110	D 4210	D 4310	D 4410

MedentiBASE® prosthetic components

- D 4840 MedentiBASE® titanium base incl. bridge screw
- D 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	D 4840	D 7600

MedentiBASE® adhesive cap for passive-fit

- D 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- D 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	D 4800	D 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	D 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» E-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The E-Series abutments are compatible with

Nobel Biocare NobelReplace® Tapered


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over cam

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	NP 3,5	RP 4,3	WP 5,0	D 6,0
Article No.	E 50	E 51	E 52	E 53

Abutment screw

- Titanium grade 5 KV
- UG



Article No.		E 60	E 61
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.		

Laboratory screw

- red coated



Article No.		E 60 L	E 61 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.		

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,3	WP 5,0	D 6,0
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm	0,3 mm
Article No.	E 1000	E 1010	E 1020	E 1030
Abutment screw Laboratory screw	E 60 E 60 L	E 61 E 61 L	E 61 E 61 L	E 61 E 61 L
Please note:	This is used with Scanbody 2. generation.			

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,3	WP 5,0	D 6,0
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm	0,3 mm
Article No.	E 1100	E 1110	E 1120	E 1130
Abutment screw Laboratory screw	E 60 E 60 L	E 61 E 61 L	E 61 E 61 L	E 61 E 61 L
Please note:	This is used with Scanbody 2. generation.			

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,3	WP 5,0
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	E 1500	E 1510	E 1520
Abutment screw Laboratory screw	E 60 E 60 L	E 61 E 61 L	E 61 E 61 L
Please note:	This is used with Scanbody 2. generation.		

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,3	WP 5,0
Diameter	11.5 mm	11.5 mm	11.5 mm
Article No.	E 9000-AG	E 9010-AG	E 9020-AG
Abutment screw Laboratory screw	E 60 E 60 L	E 61 E 61 L	E 61 E 61 L
Please note:	This is used with Scanbody 2. generation.		

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	NP 3,5	RP 4,3	WP 5,0	D 6,0
Article No.	E 1400	E 1410	E 1420	E 1430
Please note:	The Scanbody is sterilisable and for intra-oral scanning.			

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	NP 3,5	NP 3,5	NP 3,5	NP 3,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	E 4000	E 4100	E 4200	E 4300	E 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RP 4,3	RP 4,3	RP 4,3	RP 4,3	RP 4,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	E 4010	E 4110	E 4210	E 4310	E 4410

MedentiBASE® prosthetic components

- E 4840 MedentiBASE® titanium base incl. bridge screw
- E 7600 MedentiBASE® bridge screw UG
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	E 4840	E 7600

MedentiBASE® adhesive cap for passive-fit

- E 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- E 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	E 4800	E 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	E 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

EV-SERIES



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The EV-Series abutments are compatible with

Dentsply® Implants
ASTRA TECH OsseoSpeed® EV

EV
Series

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	D 3,0	D 3,6	D 4,2	D 4,8	D 5,4
Article No.	EV 50	EV 51	EV 52	EV 53	EV 54

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	EV 60	EV 61	EV 62	EV 63
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.			

Laboratory screw

- red coated



Article No.	EV 60 L	EV 61 L	EV 62 L	EV 63 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.			

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,0	D 3,0	D 3,6	D 3,6	D 4,2
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,65 mm	1,15 mm	0,65 mm	1,15 mm	0,65 mm
Article No.	EV 1000	EV 1200	EV 1010	EV 1210	EV 1020
Abutment screw	EV 60	EV 60	EV 61	EV 61	EV 62
Laboratory screw	EV 60 L	EV 60 L	EV 61 L	EV 61 L	EV 62 L
Please note:	This is used with Scanbody 2. generation.				

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 4,2	D 4,8	D 4,8	D 5,4	D 5,4
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	1,15 mm	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	EV 1220	EV 1030	EV 1230	EV 1040	EV 1240
Abutment screw Laboratory screw	EV 62 EV 62 L	EV 63 EV 63 L	EV 63 EV 63 L	EV 63 EV 63 L	EV 63 EV 63 L
Please note:	This is used with Scanbody 2. generation.				

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,0	D 3,0	D 3,6	D 3,6	D 4,2
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,65 mm	1,15 mm	0,65 mm	1,15 mm	0,65 mm
Article No.	EV 1100	EV 1300	EV 1110	EV 1310	EV 1120
Abutment screw Laboratory screw	EV 60 EV 60 L	EV 60 EV 60 L	EV 61 EV 61 L	EV 61 EV 61 L	EV 62 EV 62 L
Please note:	This is used with Scanbody 2. generation.				

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 4,2	D 4,8	D 4,8	D 5,4	D 5,4
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	1,15 mm	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	EV 1320	EV 1130	EV 1330	EV 1140	EV 1340
Abutment screw Laboratory screw	EV 62 EV 62 L	EV 63 EV 63 L	EV 63 EV 63 L	EV 63 EV 63 L	EV 63 EV 63 L
Please note:	This is used with Scanbody 2. generation.				

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,0	D 3,6	D 4,2	D 4,8	D 5,4
Diameter	11.5 mm	11.5 mm	11.5 mm	11.5 mm	11.5 mm
Article No.	EV 9000-AG	EV 9010-AG	EV 9020-AG	EV 9030-AG	EV 9040-AG
Abutment screw Laboratory screw	EV 60 EV 60 L	EV 61 EV 61 L	EV 62 EV 62 L	EV 63 EV 63 L	EV 63 EV 63 L
Please note:	This is used with Scanbody 2. generation.				

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,0	D 3,6	D 4,2	D 4,8	D 5,4
Article No.	EV 1400	EV 1410	EV 1420	EV 1430	EV 1440
Please note:	The Scanbody is sterilisable and for intra-oral scanning.				

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» F-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

The F-Series abutments are compatible with

Nobel Biocare®
NobelActive® / NobelReplace® Conical


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	D 3,0	NP 3,5	RP 4,3/5,0
Article No.	F 52	F 50	F 51

Abutment screw

- Titanium grade 5 KV
- UG



Article No.	F 60	F 61	F 66
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.		

Laboratory screw

- red coated



Article No.	F 60 L	F 61 L	F 66 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.		

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque:
  15 Ncm: D 3,0
  35 Ncm: NP 3,5
  35 Ncm: RP 4,3/5,0



Implant connection	D 3,0	NP 3,5	NP 3,5	RP 4,3/5,0	RP 4,3/5,0
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,65 mm	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	F 1020	F 1000	F 1200	F 1010	F 1210
Abutment screw Laboratory screw	F 66 F 66 L	F 60 F 60 L	F 60 F 60 L	F 61 F 61 L	F 61 F 61 L
Please note:	This is used with Scanbody 2. generation.				

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque:
  15 Ncm: D 3,0
  35 Ncm: NP 3,5
  35 Ncm: RP 4,3/5,0



Implant connection	D 3,0	NP 3,5	NP 3,5	RP 4,3/5,0	RP 4,3/5,0
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,65 mm	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	F 1120	F 1100	F 1300	F 1110	F 1310
Abutment screw Laboratory screw	F 66 F 66 L	F 60 F 60 L	F 60 F 60 L	F 61 F 61 L	F 61 F 61 L
Please note:	This is used with Scanbody 2. generation.				

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque:
  15 Ncm: D 3,0
  35 Ncm: NP 3,5
  35 Ncm: RP 4,3/5,0



Implant connection	D 3,0	NP 3,5	RP 4,3/5,0
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,65 mm	0,65 mm	0,65 mm
Article No.	F 1520	F 1500	F 1510
Abutment screw Laboratory screw	F 66 F 66 L	F 60 F 60 L	F 61 F 61 L
Please note:	This is used with Scanbody 2. generation.		

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,3/5,0
Diameter	11.5 mm	11.5 mm
Article No.	F 9000-AG	F 9010-AG
Abutment screw Laboratory screw	F 60 F 60 L	F 61 F 61 L
Please note:	This is used with Scanbody 2. generation.	

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,0	NP 3,5	RP 4,3/5,0
Article No.	F 1420	F 1400	F 1410
Please note:	The Scanbody is sterilisable and for intra-oral scanning.		

F Series

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	NP 3,5	NP 3,5	NP 3,5	NP 3,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	F 4000	F 4100	F 4200	F 4300	F 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RP 4,3/5,0	RP 4,3/5,0	RP 4,3/5,0	RP 4,3/5,0	RP 4,3/5,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	F 4010	F 4110	F 4210	F 4310	F 4410

MedentiBASE® prosthetic components

- F 4840 MedentiBASE® titanium base incl. bridge screw
- F 7600 MedentiBASE® bridge screw UG
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	F 4840	F 7600

MedentiBASE® adhesive cap for passive-fit

- F 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- F 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	F 4800	F 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	F 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» H-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The H-Series abutments are compatible with

BIOMET 3i®
Certain®


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

• Titanium grade 5 KV



Implant connection	D 3,4	D 4,1	D 5,0
Article No.	H 50	H 51	H 52

Abutment screw

• Titanium grade 5 KV
• Hex 1.2



Article No.	H 60
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

• red coated



Article No.	H 60 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

• Titanium grade 5 KV
• incl. abutment screw
• Recommended torque: 20 Ncm



Implant connection	D 3,4	D 4,1	D 5,0
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	H 1000	H 1010	H 1020
Abutment screw Laboratory screw	H 60 H 60 L	H 60 H 60 L	H 60 H 60 L
Please note:	This is used with Scanbody 2. generation.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection		D 3,4	D 4,1	D 5,0
Chimney height		5.5 mm	5.5 mm	5.5 mm
Gingiva height		0,3 mm	0,3 mm	0,3 mm
Article No.		H 1100	H 1110	H 1120
Abutment screw Laboratory screw		H 60 H 60 L	H 60 H 60 L	H 60 H 60 L
Please note:	This is used with Scanbody 2. generation.			

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection		D 3,4	D 4,1	D 5,0
Chimney height		3.5 mm	3.5 mm	3.5 mm
Gingiva height		0,3 mm	0,3 mm	0,3 mm
Article No.		H 1500	H 1510	H 1520
Abutment screw Laboratory screw		H 60 H 60 L	H 60 H 60 L	H 60 H 60 L
Please note:	This is used with Scanbody 2. generation.			

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection		D 3,4	D 4,1	D 5,0
Diameter		11.5 mm	11.5 mm	11.5 mm
Article No.		H 9000-AG	H 9010-AG	H 9020-AG
Abutment screw Laboratory screw		H 60 H 60 L	H 60 H 60 L	H 60 H 60 L
Please note:	This is used with Scanbody 2. generation.			

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection		D 3,4	D 4,1/5,0
Article No.		H 1400	H 1410
Please note:	The Scanbody is sterilisable and for intra-oral scanning.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 3,4	D 3,4	D 3,4	D 3,4	D 3,4
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	H 4000	H 4100	H 4200	H 4300	H 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 4,1	D 4,1	D 4,1	D 4,1	D 4,1
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	H 4010	H 4110	H 4210	H 4310	H 4410

MedentiBASE® prosthetic components

- H 4840 MedentiBASE® titanium base incl. bridge screw
- H 7600 MedentiBASE® bridge screw Hex 1.2
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	H 4840	H 7600

MedentiBASE® adhesive cap for passive-fit

- H 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- H 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	H 4800	H 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	H 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» I-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

The I-Series abutments are compatible with

BIOMET 3i®
External Hex


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	D 3,4	D 4,1	D 5,0
Article No.	I 50	I 51	I 52

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	I 61
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	I 61 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	D 3,4	D 4,1	D 5,0
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,6 mm	0,6 mm	0,6 mm
Article No.	I 1000	I 1010	I 1020
Abutment screw Laboratory screw	I 61 I 61 L	I 61 I 61 L	I 61 I 61 L
Please note:	This is used with Scanbody 2. generation.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection		D 3,4	D 4,1	D 5,0
Chimney height		5.5 mm	5.5 mm	5.5 mm
Gingiva height		0,6 mm	0,6 mm	0,6 mm
Article No.		I 1100	I 1110	I 1120
Abutment screw Laboratory screw		I 61 I 61 L	I 61 I 61 L	I 61 I 61 L
Please note:	This is used with Scanbody 2. generation.			

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection		D 3,4	D 4,1	D 5,0
Chimney height		3.5 mm	3.5 mm	3.5 mm
Gingiva height		0,6 mm	0,6 mm	0,6 mm
Article No.		I 1500	I 1510	I 1520
Abutment screw Laboratory screw		I 61 I 61 L	I 61 I 61 L	I 61 I 61 L
Please note:	This is used with Scanbody 2. generation.			

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection		D 3,4	RP 4,1	D 5,0
Diameter		11.5 mm	11.5 mm	11.5 mm
Article No.		I 9000-AG	I 9010-AG	I 9020-AG
Abutment screw Laboratory screw		I 61 I 61 L	I 61 I 61 L	I 61 I 61 L
Please note:	This is used with Scanbody 2. generation.			

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection		D 3,4	D 4,1/5,0
Article No.		I 1400	I 1410
Please note:	The Scanbody is sterilisable and for intra-oral scanning.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	D 3,4	D 3,4	D 3,4
Gingiva height	2,5 mm	3,5 mm	4,5 mm
Article No.	I 4000	I 4100	I 4200

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	D 4,1	D 4,1	D 4,1
Gingiva height	2,5 mm	3,5 mm	4,5 mm
Article No.	I 4010	I 4110	I 4210

MedentiBASE® prosthetic components

- I 4840 MedentiBASE® titanium base incl. bridge screw
- I 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	I 4840	I 7600

MedentiBASE® adhesive cap for passive-fit

- I 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- I 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	I 4800	I 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	I 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» K-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The K-Series abutments are compatible with

Nobel Biocare®
Brånemark®


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	NP 3,5	RP 4,1	WP 5,1
Article No.	K 50	K 51	K 52

Abutment screw

- Titanium grade 5 KV
- UG



Article No.	K 60	K 61	K 62
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.		

Laboratory screw

- red coated



Article No.	K 60 L	K 61 L	K 62 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.		

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,1	WP 5,1
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,6 mm	0,6 mm	0,6 mm
Article No.	K 1000	K 1010	K 1020
Abutment screw Laboratory screw	K 60 K 60 L	K 61 K 61 L	K 62 K 62 L
Please note:	This is used with Scanbody 2. generation.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection		NP 3,5	RP 4,1	WP 5,1
Chimney height		5.5 mm	5.5 mm	5.5 mm
Gingiva height		0,6 mm	0,6 mm	0,6 mm
Article No.		K 1100	K 1110	K 1120
Abutment screw Laboratory screw		K 60 K 60 L	K 61 K 61 L	K 62 K 62 L
Please note:	This is used with Scanbody 2. generation.			

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection		NP 3,5	RP 4,1	WP 5,1
Chimney height		3.5 mm	3.5 mm	3.5 mm
Gingiva height		0,6 mm	0,6 mm	0,6 mm
Article No.		K 1500	K 1510	K 1520
Abutment screw Laboratory screw		K 60 K 60 L	K 61 K 61 L	K 62 K 62 L
Please note:	This is used with Scanbody 2. generation.			

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection		NP 3,5	RP 4,1	WP 5,1
Diameter		11.5 mm	11.5 mm	11.5 mm
Article No.		K 9000-AG	K 9010-AG	K 9020-AG
Abutment screw Laboratory screw		K 60 K 60 L	K 61 K 61 L	K 62 K 62 L
Please note:	This is used with Scanbody 2. generation.			

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection		NP 3,5	RP 4,1	WP 5,1
Article No.		K 1400	K 1410	K 1420
Please note:	The Scanbody is sterilisable and for intra-oral scanning.			

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	NP 3,5	NP 3,5
Gingiva height	2,5 mm	3,5 mm	4,5 mm
Article No.	K 4000	K 4100	K 4200

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RP 4,1	RP 4,1	RP 4,1
Gingiva height	2,5 mm	3,5 mm	4,5 mm
Article No.	K 4010	K 4110	K 4210

MedentiBASE® prosthetic components

- K 4840 MedentiBASE® titanium base incl. bridge screw
- K 7600 MedentiBASE® bridge screw UG
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	K 4840	K 7600

MedentiBASE® adhesive cap for passive-fit

- K 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- K 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	K 4800	K 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	K 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» L-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

The L-Series abutments are compatible with

Straumann®
Bone Level®


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over edge

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV




	NC 3,3	RC 4,1/4,8
Implant connection		
Article No.	L 50	L 51

Abutment screw

- Titanium grade 5 KV
- Torx T6




Article No.	L 60	L 63
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.	

Laboratory screw

- red coated



Article No.	L 60 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm




		NC 3,3	RC 4,1/4,8
Implant connection			
Chimney height		3.5 mm	3.5 mm
Gingiva height		1,0 mm	0,8 mm
Article No.		L 1000	L 1010
Abutment screw Laboratory screw		L 60 L 60 L	L 60 L 60 L
Please note:	This is used with Scanbody 2. generation.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection		NC 3,3	RC 4,1/4,8
Chimney height		5.5 mm	5.5 mm
Gingiva height		1,0 mm	0,8 mm
Article No.		L 1100	L 1110
Abutment screw Laboratory screw		L 60 L 60 L	L 60 L 60 L
Please note:	This is used with Scanbody 2. generation.		

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection		NC 3,3	RC 4,1/4,8
Chimney height		3.5 mm	3.5 mm
Gingiva height		1,0 mm	0,8 mm
Article No.		L 1500	L 1510
Abutment screw Laboratory screw		L 60 L 60 L	L 63
Please note:	This is used with Scanbody 2. generation.		

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection		NC 3,3	RC 4,1/4,8
Diameter		11.5 mm	11.5 mm
Article No.		L 9000-AG	L 9010-AG
Abutment screw Laboratory screw		L 60 L 60 L	L 60 L 60 L
Please note:	This is used with Scanbody 2. generation.		

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection		NC 3,3	RC 4,1/4,8
Article No.		L 1400	L 1410
Please note:	The Scanbody is sterilisable and for intra-oral scanning.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NC 3,3	NC 3,3	NC 3,3	NC 3,3	NC 3,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	L 4000	L 4100	L 4200	L 4300	L 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RC 4,1/4,8	RC 4,1/4,8	RC 4,1/4,8	RC 4,1/4,8	RC 4,1/4,8
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	L 4010	L 4110	L 4210	L 4310	L 4410

MedentiBASE® prosthetic components

- L 4840 MedentiBASE® titanium base incl. bridge screw
- L 7600 MedentiBASE® bridge screw Torx T6
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	L 4840	L 7600

MedentiBASE® adhesive cap for passive-fit

- L 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- L 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	L 4800	L 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	L 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» N-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

The N-Series abutments are compatible with

Straumann®
Tissue Level®

DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Article No.	N 53	N 51	N 52

Abutment screw

- Titanium grade 5 KV
- Torx T6



Article No.	N 62	N 65
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.	

Laboratory screw

- red coated



Article No.	N 62 L	N 65 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.	

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,6 mm	0,1 mm	0,1 mm
Article No.	N 1030	N 1010	N 1020
Abutment screw Laboratory screw	N 65 N 65 L	N 62 N 62 L	N 62 N 62 L
Please note:	This is used with Scanbody 2. generation.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Chimney height	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,6 mm	0,1 mm	0,1 mm
Article No.	N 1130	N 1110	N 1120
Abutment screw Laboratory screw	N 65 N 65 L	N 62 N 62 L	N 62 N 62 L
Please note:	This is used with Scanbody 2. generation.		

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,35 mm	0,1 mm	0,1 mm
Article No.	N 1530	N 1510	N 1520
Abutment screw Laboratory screw	N 65 N 65 L	N 62 N 62 L	N 62 N 62 L
Please note:	This is used with Scanbody 2. generation.		

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Diameter	11.5 mm	11.5 mm	11.5 mm
Article No.	N 9030-AG	N 9010-AG	N 9020-AG
Abutment screw Laboratory screw	N 65 N 65 L	N 62 N 62 L	N 62 N 62 L
Please note:	This is used with Scanbody 2. generation.		

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Article No.	N 1430	N 1410	N 1420
Please note:	The Scanbody is sterilisable and for intra-oral scanning.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	NNC 3,5	NNC 3,5	NNC 3,5	NNC 3,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	N 4030	N 4130	N 4230	N 4330	N 4430

MedentiBASE® prosthetic components

- NNC 4840 MedentiBASE® titanium base incl. bridge screw
- NNC 7600 MedentiBASE® bridge screw Torx T6
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	NNC 4840	NNC 7600

MedentiBASE® adhesive cap for passive-fit

- NNC 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- NNC 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	NNC 4800	NNC 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw
- sterilisable and for intra-oral scanning



Article No.	NNC 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RN 4,8
Gingiva height	0 mm
Article No.	N 4010

MedentiBASE® prosthetic components

- N 7600 MedentiBASE® bridge screw Torx T6
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV
Article No.	N 7600

MedentiBASE® adhesive cap for passive-fit

- N 4800 MedentiBASE adhesive cap Short incl. bridge screw



Version	Short
Article No.	N 4800

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	N 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment
- N 51 Laboratory implant





Article No.	M 11-6	N 51

N Series

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

R-SERIES



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The R-Series abutments are compatible with

Zimmer Dental®
- Tapered Screw-Vent®

MIS®
- SEVEN® Internal Hex

BioHorizons®
- Tapered Internal
- Internal Plus
- Tapered Tissue Level


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Comparison chart

Zimmer Dental®
MIS®
BioHorizons®

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Zimmer Dental®	Platform 3,5 mm		Platform 4,5 mm	Platform 5,7 mm
Implant Ø	**3,70**	**4,10**	**4,70**	**6,00**
Laboratory implant Ø	3,50	3,50	4,50	5,70
Article Number	IA3	IA3	IA4	IA5
Implant connection				
Medentika R-Series Laboratory implant	**3,5**	**3,5**	**4,5**	**5,7**
Article Number	R 50	R 50	R 51	R 52

MIS®	Narrow-Platform		Standard-Platform		Wide-Platform	
Implant Ø	**3,30**	**3,30**	**3,75**	**4,20**	**5,00**	**6,00**
Laboratory implant Ø	3,30	4,20	4,20	4,20	4,50	4,50
Article Number	MN-RSM10	MD-RSM10	MD-RSM10	MD-RSM10	MW-RSM10	MW-RSM10
Implant connection						
Medentika R-Series Laboratory implant	**XX**	**3,50**	**3,50**	**3,50**	**4,50**	**4,50**
Article Number	XX	R 50	R 50	R 50	R 51	R 51

BioHorizons®	Platform 3,0 mm	Platform 3,5 mm	Platform 4,5 mm	Platform 5,7 mm
Implant Ø	**3,00**	**3,80**	**4,60**	**5,80**
Laboratory implant Ø	3,00	3,50	4,50	5,70
Article Number	TP3IA	SYIA	SGIA	SBIA
Implant connection				
Medentika R-Series Laboratory implant	**XX**	**3,50**	**4,50**	**5,70**
Article Number	XX	R 50	R 51	R 52

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	D 3,5	D 4,5	D 5,7
Article No.	R 50	R 51	R 52

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	R 60
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	R 60 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,5	D 4,5	D 5,7
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	R 1000	R 1010	R 1020
Abutment screw Laboratory screw	R 60 R 60 L	R 60 R 60 L	R 60 R 60 L
Please note:	This is used with Scanbody 2. generation.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection		D 3,5	D 4,5	D 5,7
Chimney height		5.5 mm	5.5 mm	5.5 mm
Gingiva height		0,3 mm	0,3 mm	0,3 mm
Article No.		R 1100	R 1110	R 1120
Abutment screw Laboratory screw		R 60 R 60 L	R 60 R 60 L	R 60 R 60 L
Please note:	This is used with Scanbody 2. generation.			

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection		D 3,5	D 4,5	D 5,7
Chimney height		3.5 mm	3.5 mm	3.5 mm
Gingiva height		0,3 mm	0,3 mm	0,3 mm
Article No.		R 1500	R 1510	R 1520
Abutment screw Laboratory screw		R 60 R 60 L	R 60 R 60 L	R 60 R 60 L
Please note:	This is used with Scanbody 2. generation.			

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection		D 3,5	D 4,5	D 5,7
Diameter		11.5 mm	11.5 mm	11.5 mm
Article No.		R 9000-AG	R 9010-AG	R 9020-AG
Abutment screw Laboratory screw		R 60 R 60 L	R 60 R 60 L	R 60 R 60 L
Please note:	This is used with Scanbody 2. generation.			

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection		D 3,5	D 4,5	D 5,7
Article No.		R 1400	R 1410	R 1420
Please note:	The Scanbody is sterilisable and for intra-oral scanning.			

R Series

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 30 Ncm



Implant connection	D 3,5	D 3,5	D 3,5	D 3,5	D 3,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	R 4000	R 4100	R 4200	R 4300	R 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 30 Ncm



Implant connection	D 4,5	D 4,5	D 4,5	D 4,5	D 4,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	R 4010	R 4110	R 4210	R 4310	R 4410

MedentiBASE® prosthetic components

- R 4840 MedentiBASE® titanium base incl. bridge screw
- R 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	R 4840	R 7600

MedentiBASE® adhesive cap for passive-fit

- R 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- R 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	R 4800	R 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	R 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

S-SERIES



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The S-Series abutments are compatible with

DENTSPLY Implants®
ASTRA TECH OsseoSpeed® TX

S
Series

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Titanium grade 5 KV



Implant connection	D 3,0	D 3,5/4,0	D 4,5/5,0
Article No.	S 53	S 50	S 52

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	S 60	S 61	S 67
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.		

Laboratory screw

- red coated



Article No.	S 60 L	S 61 L	S 67 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.		

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque:
  15 Ncm: D 3,0
  25 Ncm: D 3,5/4,0
  25 Ncm: D 4,5/5,0



Implant connection	D 3,0	D 3,5/4,0	D 3,5/4,0	D 4,5/5,0	D 4,5/5,0
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,6 mm	0,6 mm	1,1 mm	0,6 mm	1,1 mm
Article No.	S 1030	S 1000	S 1200	S 1020	S 1220
Abutment screw Laboratory screw	S 67 S 67 L	S 60 S 60 L	S 60 S 60 L	S 61 S 61 L	S 61 S 61 L
Please note:	This is used with Scanbody 2. generation.				

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque:
  15 Ncm: D 3,0
  25 Ncm: D 3,5/4,0
  25 Ncm: D 4,5/5,0



Implant connection	D 3,0	D 3,5/4,0	D 3,5/4,0	D 4,5/5,0	D 4,5/5,0
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,6 mm	0,6 mm	1,1 mm	0,6 mm	1,1 mm
Article No.	S 1130	S 1100	S 1300	S 1120	S 1320
Abutment screw Laboratory screw	S 67 S 67 L	S 60 S 60 L	S 60 S 60 L	S 61 S 61 L	S 61 S 61 L
Please note:	This is used with Scanbody 2. generation.				

Titanium base Bridges/ bars 2nd Generation


- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection				D 3,5/4,0	D 4,5/5,0
Chimney height				3.5 mm	3.5 mm
Gingiva height				0,6 mm	0,6 mm
Article No.				S 1500	S 1520
Abutment screw Laboratory screw				S 60 S 60 L	S 61 S 61 L
Please note:	This is used with Scanbody 2. generation.				

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection				D 3,5/4,0	D 4,5/5,0
Diameter				11.5 mm	11.5 mm
Article No.				S 9000-AG	S 9020-AG
Abutment screw Laboratory screw				S 60 S 60 L	S 61 S 61 L
Please note:	This is used with Scanbody 2. generation.				

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,0	D 3,5/4,0	D 4,5/5,0
Article No.	S 1430	S 1400	S 1420
Please note:	The Scanbody is sterilisable and for intra-oral scanning.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 3,5/4,0	D 3,5/4,0	D 3,5/4,0	D 3,5/4,0	D 3,5/4,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	S 4000	S 4100	S 4200	S 4300	S 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 4,5/5,0	D 4,5/5,0	D 4,5/5,0	D 4,5/5,0	D 4,5/5,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	S 4020	S 4120	S 4220	S 4320	S 4420

MedentiBASE® prosthetic components

- S 4840 MedentiBASE® titanium base incl. bridge screw
- S 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	S 4840	S 7600

MedentiBASE® adhesive cap for passive-fit

- S 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- S 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	S 4800	S 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	S 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» T-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The T-Series abutments are compatible with

DENTSPLY Implants®
XiVE® S


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

T Series

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Stainless steel



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Article No.	T 50	T 55	T 51	T 52

Abutment screw

- Titanium grade 5 KV
- Hex 1.2



Article No.	T 60
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	T 60 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,6 mm	0,6 mm
Article No.	T 1000	T 1005	T 1010	T 1020
Abutment screw Laboratory screw	T 60 T 60 L	T 60 T 60 L	T 60 T 60 L	T 60 T 60 L
Please note:	This is used with Scanbody 2. generation.			

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,3 mm	0,3 mm	0,6 mm	0,6 mm
Article No.	T 1100	T 1105	T 1110	T 1120
Abutment screw Laboratory screw	T 60 T 60 L	T 60 T 60 L	T 60 T 60 L	T 60 T 60 L
Please note:	This is used with Scanbody 2. generation.			

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,6 mm	0,6 mm
Article No.	T 1500	T 1505	T 1510	T 1520
Abutment screw Laboratory screw	T 60 T 60 L	T 60 T 60 L	T 60 T 60 L	T 60 T 60 L
Please note:	This is used with Scanbody 2. generation.			

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Diameter	11.5 mm	11.5 mm	11.5 mm	11.5 mm
Article No.	T 9000-AG	T 9005-AG	T 9010-AG	T 9020-AG
Abutment screw Laboratory screw	T 60 T 60 L	T 60 T 60 L	T 60 T 60 L	T 60 T 60 L
Please note:	This is used with Scanbody 2. generation.			

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,4	D 3,8	D 4,5/5,5
Article No.	T 1400	T 1405	T 1410
Please note:	The Scanbody is sterilisable and for intra-oral scanning.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,4	D 3,4	D 3,4	D 3,4
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	T 4000	T 4100	T 4200	T 4300	T 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 3,8	D 3,8	D 3,8	D 3,8	D 3,8
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	T 4005	T 4105	T 4205	T 4305	T 4405

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 4,5	D 4,5	D 4,5	D 4,5	D 4,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	T 4010	T 4110	T 4210	T 4310	T 4410

MedentiBASE® prosthetic components

- T 4840 MedentiBASE® titanium base incl. bridge screw
- T 7600 MedentiBASE® bridge screw Hex 1.2
- Recommended torque: 15 Ncm



Material				Titanium grade 5 KV	Titanium grade 5 KV
Article No.				T 4840	T 7600

MedentiBASE® adhesive cap for passive-fit

- T 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- T 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version				Short	Long
Article No.				T 4800	T 4820

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	T 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» Y-SERIES «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

The Y-Series abutments are compatible with

DENTSPLY Implants®
ANKYLOS®

Products indicated with ® are registered brand names of respective manufactures.

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Laboratory implant

- Stainless steel



Implant connection	D 3,5-7,0
Article No.	Y 50

Abutment screw

- Titanium grade 5 KV
- Hex 1.0



Article No.	Y 61
Please note:	The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw


- red coated



Article No.	Y 61 L
Please note:	The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection		D 3,5-7,0	D 3,5-7,0
Chimney height		3.5 mm	3.5 mm
Gingiva height		0,65 mm	1,15 mm
Article No.		Y 1500	Y 1700
Abutment screw Laboratory screw		Y 61 Y 61 L	Y 61 Y 61 L
Please note:	This is used with Scanbody 2. generation.		

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Titanium base Zirconium abutment 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5-7,0	D 3,5-7,0
Chimney height	5.5 mm	5.5 mm
Gingiva height	0,65 mm	1,15 mm
Article No.	Y 1600	Y 1800
Abutment screw Laboratory screw	Y 61 Y 61 L	Y 61 Y 61 L

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection		D 3,5-7,0
Article No.		Y 1400
Please note:	The Scanbody is sterilisable and for intra-oral scanning.	

MedentiBASE abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 3,5-7,0	D 3,5-7,0	D 3,5-7,0	D 3,5-7,0	D 3,5-7,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	Y 4000	Y 4100	Y 4200	Y 4300	Y 4400

MedentiBASE prosthetic components

- Y 4840 MedentiBASE® titanium base incl. bridge screw
- Y 7600 MedentiBASE® bridge screw Hex 1.0
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	Y 4840	Y 7600

MedentiBASE adhesive cap for passive-fit

- Y 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- Y 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	Y 4800	Y 4820

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	Y 4900
Please note:	The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

» Tools «



https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

Placement instrument MedentiBASE® abutment

- Hardened stainless steel



Version	Manual and ratchet
Type	2-piece
Article No.	M 11-6

Torque ratchet

- with infinitely variable torque setting
- up to max. 40 Ncm
- Hardened stainless steel



Article No.	M 12

Wooden tool box Placement instrument Manual and ratchet



Version	long
Article No.	MB 04K
Content:	Hex 1,26 / Hex 1,2 / Hex 1,0 / Torx T6 / UG

Compartment box Abutments and scanbodys

- 30 compartments
- Polystyrene



Version	without contents
Article No.	MB 04

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ

stomdevice

AMANNGIRRBACH

EN
IMPLANT PROSTHETICS
9IMPME EN 2017 03
heiblau.com

HERSTELLER:

MEDENTIKA GMBH
Hammerweg 8-10
76549 Hügelsheim
Fon +49 7229 69912-0
Fax +49 7229 69912-20
info@medentika.de
www.medentika.de

VERTRIEB:

AUSTRIA (HEADQUARTERS)
Amann Girrbach AG
Koblach, Austria
Fon +43 5523 62333-105
austria@amanngirrbach.com

GERMANY
Amann Girrbach GmbH
Pforzheim, Germany
Fon +49 7231 957-100
germany@amanngirrbach.com

NORTH AMERICA
Amann Girrbach North America, LP
Charlotte, NC, U.S.A.
Fon +1 704 837 1404
america@amanngirrbach.com

BRASIL
Amann Girrbach Brasil LTDA
Curitiba, Brasil
Fon +55 41 3287 0897
brasil@amanngirrbach.com

ASIA
Amann Girrbach Asia PTE LTD.
Singapore, Asia
Fon +65 6592 5190
singapore@amanngirrbach.com

CHINA
Amann Girrbach China Co., Ltd.
Beijing, China
Fon +86 10 8886 6064
china@amanngirrbach.com

WWW.AMANNGIRRBACH.COM

AMANNGIRRBACH

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ