stomdevice

INSTRUCTION FOR USE

Nexcomp Nano Hybrid Composite Resin

Anterior and posterior light cured restorative

Nexcomp is Nano Hybrid Composite Resin for Anterior and Posterior Resorations. Nexcomp is reinforced nanofiller composite

Nexcomp is the ideal nanofilled microhybrid to use both anteriorly and posteriorly. Leading clinicians agree that this versatile material is in a class by itself

INDICATION

- Class Ⅰ, Ⅱ, Ⅲ, Ⅳ, Ⅴ restorations
- Core-buildup to replace missing tooth structure
- Diastema closures
- Direct veneers
- Composite and porcelain repairs

CONTRAINDICATIONS

- **Nexcomp** should not be used if patients are known to be allergic to methacrylate resins.

INSTRUCTIONS

① Select the desired type and shade of **Nexcomp** composite prior to isolating and preparing
② Clean the preparation with a slurry of pumice and water. Rinse and dry the preparation
③ Apply 37% Meta Etchant gel to available enamel / dentin for 15 seconds. Remove gel and rinse for 20 seconds, Blot dry keep moist, do not rub.
④ Apply Bond to the tooth-Apply a double coat of Meta P&Bond by using mircrobrush. Gently air dry for 5 seconds. Light cure for 10 seconds. If you apply other product to saturate all internal surfaces, or bonding agent according to manufacturer's instructions
⑤ Place composite-Place **Nexcomp** in increments of 2mm or less in Anterior and posterior restorations. Cure for 20 seconds in increments of 2mm.
⑥ Finish and Polish-After contouring composite, polish surface with finishing and polishing proximal areas.
⑦ Occlusion-Check lateral and centric occlusion. Adjust as nessary.

SIDE EFFECTS

- None known to date

CAUTIONS

- Avoid contact of **Nexcomp** with the skin/mucous membrane or eyes
- None polymerized **Nexcomp** may be causes lightly, reversible irritation or lead to a general sensitization against methacrylates.
- Keep away from child.
- For dental use only
- Do not use after the date of expiration.
- Exposure to light may be lead to premature polymerization

STORAGE

- Storage 18~24℃(64~75℉)
- Close syringe cap immediately after use.
- Avoid from directly sun light.

SHELF LIFE

- 2 years from the manufacturing date

COMPONENTS

Nexcomp Set
- 4g x 7 Syringe (A1, A2, A3, A3.5, B2, A2E, B2E)

Nexcomp Kit
- 4g x 7 Syringe (A1, A2, A3, A3.5, B2, A2E, B2E)
- Meta P&Bond 5g
- Meta Etchant 3g

Nexcomp Refill
- 4g x 1 Syringe (A1B, A2B, A3B, A3.5B, B2B, A2E, B2E, A1, A2, A3, A3.5, A4, B1, B2, B3, C2, C3, D2, D3, TL, WT, OP, OA2, OA3)

SYMBOLS

- Use By
- LOT Batch Code
- CE 0120 CE-Mark and identification number of notified body. Product conforms to the essential requirements of the Medical Device Directive 93/42/EEC
- Attention, See instruction for use
- Manufacturer

Made in Korea

Manufactured by
META BIOMED CO.LTD.
634, Yeonje-ri, Gang-oe-myeon, Cheong-won-gun Chungcheongbuk-do, Korea
Tel:+82-43-216-0433
Fax:+82-43-217-1988
E-mail: info@meta-biomed.com
Web site: http://www.meta-biomed.com

META® BIOMED

DM-40-6

META BIOMED CO.LTD.

CHECKING SHEET

Nexcomp

TITLE	
CODE	DM-40-6
SIZE	207x174mm
PAPER	60g 모조
MAIN COLOR	INDIGO 2767C 별색 단도
SPECIAL	
FUNCTION	
PRINT COLOR	

INDIGO 2767C

DATE	2009. 07. 10.
DESIGNER	진희남
RECOGNITION	김상일
CONFIRMATION	META BIOMED CO.LTD.

DATE NAME SIGN

2011. 06. 02 → DM-40-6 / 주소변경 (홍선미)

https://stomdevice.ru - стоматологическое и зуботехническое оборудование с доставкой по России и СНГ